

IFB #20026 entitled “Biological Safety Cabinet and HEPA Exhaust Filter Certification”

Questions and Answers

	Category	Bidder’s Question	Answer
1.	Attachment B – Bid Form	ENV is a reseller of parts and therefore does not stock parts. Are you able to modify the bid form to include a line for parts pricing or must parts be included in cost?	No modification of the bid form is allowed. Pricing under Category D of attachment B should be all inclusive for the five maintenance and repair items listed.
2.	Minimum Qualifications to Bid, Section 3, Page 5	Do you require an NSF accredited technician or can you accept technicians who have been trained to NSF specifications and/or previously withheld an NSF accreditation?	An NSF accredited field technician will be required for all HEPA filter and biosafety cabinet certifications.
3.	Detailed Specifications, Section 4, Page 6, Paragraph 9	Can you please clarify the disposal process of the used filters?	Disposal of any filters removed from equipment or HVAC filter housings will be the responsibility of the Wadsworth Center.
4.	Detailed Specifications, Section 4, Page 6, Paragraph 11	Will an NSF accredited technician be required to make all repairs and decontaminations?	No. An NSF accredited technician will not be required to make repairs or perform decontaminations, but an NSF accredited field technician will be required to certify equipment after any repairs or decontaminations that would require re-certification.