

Invitation for Bids

C040463

Chemical Fume Hood, Biological Safety Cabinet, and HEPA Exhaust Filter Certification

Questions and Answers

QUESTION NUMBER	QUESTION	ANSWER
1	I don't see any specs on the units being requested (accepted manufacturers, unit size/dimensions, accessories required, etc.). Can you please provide a detailed spec for all units as well as all required accessories. If you don't have a spec but have information on the sizes and accessories, would you like us to provide you with our standard spec?	We are not requesting that the bidder provide us with any units (Biosafety Cabinets or chemical fume hoods). We are requesting the services in attachment B to be performed on units we already have in service or units that we may purchase independently in the future. For biosafety cabinets we primarily use Baker, but we do have cabinets from multiple other manufacturers. Manufacturers of in use fume hoods is more variable. Bedding dump stations are primarily NuAire and microisolators are primarily Techniplast.
2	I see the qty being requested which are noted as "yearly qty", can you confirm how many total years this project will consist of?	The numbers given in the "yearly quantity" column of attachment B indicate the number of times we reasonably anticipate that particular service to be performed per fiscal year. The contract will be awarded for 5 years.